AMENDMENT OF SOLICITATION/MODIFICATION OF CONTRACT

1. CONTRACT ID CODE	PAGE	OF	PAGES
	1		3

2. AMENDMENT/MODIFICATION NUMBER	3. EFFECTIVE DATE	4. REQUISITION/PURCHASE REQUISITION NUMBER	5. PROJECT NUMBER *(If applicable)*
0001	06/25/2026		

6. ISSUED BY CODE 70FA20

FEDERAL EMERGENCY MANAGEMENT AGENCY
OCPO PREPAREDNESS SECTION NETC
16825 SOUTH SETON AVENUE
EMMITSBURG MD 21727

7. ADMINISTERED BY *(If other than Item 6)* CODE

8. NAME AND ADDRESS OF CONTRACTOR *(Number, street, county, State and ZIP Code)*

CODE | FACILITY CODE

(X) 9A. AMENDMENT OF SOLICITATION NUMBER
70FA4026Q00000014

[X] 9B. DATED *(SEE ITEM 11)*
06/25/2026

[ ] 10A. MODIFICATION OF CONTRACT/ORDER NUMBER

10B. DATED *(SEE ITEM 13)*

11. THIS ITEM ONLY APPLIES TO AMENDMENTS OF SOLICITATIONS

[X] The above numbered solicitation is amended as set forth in Item 14. The hour and date specified for receipt of Offers [ ] is extended. [X] is not extended.

Offers must acknowledge receipt of this amendment prior to the hour and date specified in the solicitation or as amended, by one of the following methods: (a) By completing items 8 and 15, and returning _____ copies of the amendment; (b) By acknowledging receipt of this amendment on each copy of the offer submitted; or (c) By separate letter or electronic communication which includes a reference to the solicitation and amendment numbers. FAILURE OF YOUR ACKNOWLEDGMENT TO BE RECEIVED AT THE PLACE DESIGNATED FOR THE RECEIPT OF OFFERS PRIOR TO THE HOUR AND DATE SPECIFIED MAY RESULT IN REJECTION OF YOUR OFFER. If by virtue of this amendment you desire to change an offer already submitted, such change may be made by letter or electronic communication, provided each letter or electronic communication makes reference to the solicitation and this amendment, and is received prior to the opening hour and date specified.

12. ACCOUNTING AND APPROPRIATION DATA *(If required)*

13. THIS ITEM APPLIES ONLY TO MODIFICATIONS OF CONTRACTS/ORDERS. IT MODIFIES THE CONTRACT/ORDER NUMBER AS DESCRIBED IN ITEM 14.

CHECK ONE	
[ ]	A. THIS CHANGE ORDER IS ISSUED PURSUANT TO: *(Specify authority)* THE CHANGES SET FORTH IN ITEM 14 ARE MADE IN THE CONTRACT ORDER NUMBER IN ITEM 10A.
[ ]	B. THE ABOVE NUMBERED CONTRACT/ORDER IS MODIFIED TO REFLECT THE ADMINISTRATIVE CHANGES *(such as changes in paying office, appropriation data, etc.)* SET FORTH IN ITEM 14, PURSUANT TO THE AUTHORITY OF FAR 43.103(b).
[ ]	C. THIS SUPPLEMENTAL AGREEMENT IS ENTERED INTO PURSUANT TO AUTHORITY OF:
[ ]	D. OTHER *(Specify type of modification and authority)*

E. IMPORTANT: Contractor [ ] is not [ ] is required to sign this document and return _____ copies to the issuing office.

14. DESCRIPTION OF AMENDMENT/MODIFICATION (*Organized by UCF section headings, including solicitation/contract subject matter where feasible.)*

See continuation page

Except as provided herein, all terms and conditions of the document referenced in Item 9A or 10A, as heretofore changed, remains unchanged and in full force and effect.

15A. NAME AND TITLE OF SIGNER *(Type or print)*		16A. NAME AND TITLE OF CONTRACTING OFFICER *(Type or print)*	
		Jeffery Byrd	
15B. CONTRACTOR/OFFEROR	15C. DATE SIGNED	16B. UNITED STATES OF AMERICA	16C. DATE SIGNED
(Signature of person authorized to sign)		*(Signature of Contracting Officer)*	

NAME OF OFFEROR OR CONTRACTOR

ITEM NO. (A)	SUPPLIES/SERVICES (B)	QUANTITY (C)	UNIT (D)	UNIT PRICE (E)	AMOUNT (F)
	The purpose of this amendment is to correct the language in section 2.1.2 of the SOW, which is changed from: Exam Table with pneumatic back and electrical outlet (Same or equivocal to Brewer Access Exam Table (Gunmetal) SKU 5000 02. Change to: Exam Table with pneumatic back and electrical outlet (Same or EQUIVALENT to Brewer Access Exam Table (Gunmetal) SKU 5000 02. Also, respond to received questions. All other terms and conditions remain unchanged.				

2.1.2 The Contractor shall be responsible to deliver the patient examination table to the FEMA Occupational Health Clinic: FEMA HQ 500 C St SW, Room 7NW-0904 Washington DC 20472.

Table 1. Occupational Health Clinic Patient Examination Table	
Exam Table with pneumatic back and electrical outlet (Same or **equivalent** to Brewer Access Exam Table (Gunmetal) SKU 5000-02	• 32”h x 28” wide x 57.3” extends to 71”. • Seamless upholstery. • Large patient step, four leg step design. Step surface 17.7” long x 20.4” wide x 8” high • Pneumatic back rest for patient positioning. Multi-density, bolstered support upholstery. • Leg extension. Length 15” • Adjustable ergonomic stirrups stored in the table and manually extended through infinite positions and four lateral positions. • Pass-through side drawers accessible from right or left side of the table. Drawer stops to allow the drawer to open from only one side if desired. • Front drawer 19.8” x 19.2” x 3.8” • Side drawers 22.7” x 20.2” x 7” • Additional paper storage under backrest for up to 6 additional paper rolls. Paper roll storage capacity 23” x 23” x 3.25”. Slide through paper roll holder accepts 21” x 3” rolls.
Warranty	5 year warranty
Electrical	Hospital-grade duplex electrical outlets. 115VAC, 60Hz, 5.8 amps max. Outlets can be put on right or left.
Standards	Must conform to US electrical and medical industry codes and regulations and meet or exceed industry requirements. UL 60601 certified.