MEDPIPE

Corporate Office
P.O. Box 541
Lawrence, PA 15055
(724) 745-1854
(724) 745-2184 (fax)

Columbus Office
865 Eastwind Drive
Westerville, OH 43081
(614) 392-2222
(614) 392-2300 (fax)

Customer Information		System Information		Job Information	
Facility	VA Medical Center Erie	System Description	RAMVAC dental air duplex	Job #	622374
Address	135 East 38th Street	System Model #	OSP25	SO #	622374
	Erie,PA 16504	System Serial #	4773030	Trip #	1
Contact	Davvid Dombkowski	Manufacture Date	07/01/19	Year	2024
Email	david.dombkowski@va.org	System Location	Dental Mechanical room	Date	12/11/24
Phone	814-860-2924	Customer System #		Tech	RG/TP

Pump #	#1	#2	#3	#4
Model #	OF1202	OF1202		
Serial #	Gast Item#004822	Gast Item#004822		
Run Hours				
Room Temperature	77°F			
Discharge Temperature				
Discharge Check Valves	A	A		
Unloaders and Solenoids	A	A		
Inlet Filters	A	A		
After Cooler Drain Valves	A	A		
High Temp Shutdown				
Crankcase Filter				
Rider and Piston Rings	A	A		
Starter Contacts	A	A		
Thermal Overload Settings				
Transformer Condition				
Panel Lamps	A			
Relief Valves	A			
Panel Relays	A			
Lead Pressure Switch	80-100 psi	Local Lag Alarm		
Lag 1 Pressure Switch	psi	Audible		
Lag 2 Pressure Switch	psi	Visual		
Lag 3 Pressure Switch	psi			

LEGEND
A = Acceptable
U = Unacceptable
R = Repaired
NR = Needs Replaced
NA = Not Applicable

Motor Information Section		Motor #1	Motor # 2	Motor # 3	Motor #4
Volts L1 to L2	208	9.50 AMPS	9.30 AMPS		
Volts L2 to L3	208	9.40 AMPS	9.50 AMPS		
Volts L1 to L3	208	9.40 AMPS	9.70 AMPS		
Efficiency		Service Factor		Horse Power	RPM 1,650
Frame Size		Voltage		Full Load Amps	

CO Monitor				Dew Point Monitor	
Manuf/Model/SN:	None			Manuf/Model/SN:	None
Test Gas Reading		Gas Conc.=	20 PPM	D.P. Reading (as found)	
Cal Performed ?		Fail		Measured D.P. (Medpipe)	
Sensor Replaced ?				Sensor Replaced ?	
Reading after Cal				D.P. Reading (as left)	

Dryers					
Unit On Line (as found)		Pre Filters			Lamps	
Unit On Line (as left)		After Filters			Suction Pressure	
Selector Switch		Discharge Check Valve			Blow Down Valve	
Purge Select		Mufflers			Auto Blow Down	
Line Regulator		Solenoids			Float Drain	

MEDPIPE

Corporate Office
P.O. Box 541
Lawrence, PA 15055
(724) 745-1854
(724) 745-2184 (fax)

Columbus Office
865 Eastwind Drive
Westerville, OH 43081
(614) 392-2222
(614) 392-2300 (fax)

Parts Used		
Quantity	**Part Number**	**Part Description**

Deficiencies
Customer installed additional compressor system

Comments
customer servicing units for dental due to patient backlog

Recommendations