Veterinary Hematology Analyzer Specifications

Minimum required salient characteristics:

1. Complete Blood Count, with minimum 5-part CBC differential

2. Multiple animal species

3. Able to process micro samples of blood: 15-30 uL

4. Easy-to-use touchscreen interface

5. Automatic analysis and test results on report

6. Easy to manage reagents