## Salient Characteristics

Commercial Availability Certification:
I have evaluated the commercial marketplace and determined that all of the salient characteristics documented below are available in products/services offered in the commercial marketplace.

Signature (typed name): [redacted]
Technical Point of Contact (TPOC)

*The Government requires three (3) environmental temperature chambers with following minimum characteristics:*

1. An environmental temperature chamber that can operate continuously at -65°C to +180ºC with the ability to control temperature to +/- 1°C.
2. An environmental temperature chamber with a chamber that is greater than (>) one (1) cubic foot in volume and is at least 11 inches deep and 11 inches high.
3. An environmental chamber that operates on 120 volts (V) alternating current (AC) voltage on a 20A circuit.
4. An environmental chamber with a feedthrough port on one side for cables with a diameter of at least 2 ½ inches.
5. An upright chamber on wheels with the chamber on top and compressor underneath.
6. A chamber interior made from stainless steel.
7. A chamber capable of holding at least two (2) shelves.

Contracting Officer:

| |
|---|
| 7/30/2026<br>X Swati Jain |
| Contracting Officer (KO)<br>Signed by: USA |